

एसजेवीएन अरुण-३
पावर डेवलपमेण्ट कम्पनी (प्रा.) लि.
(भारत सरकार र हिमाचल सरकारको संयुक्त कम्पनी,
एसजेवीएन लिमिटेड बाट स्थापित कम्पनी)



SJVN Arun-3
Power Development Company (P.) Ltd.
(A company promoted by SJVN limited,
joint venture of Govt. of India and Govt. of H.P.)

Ref. No.: SAPDC/P&C/PPR-94/2019-232/1

Dated: 22.11.2019

ARUN-3 HYDROELECTRIC PROJECT (900 MW)

CLARIFICATIONS-I

Subject: PPR 94 “Supply, transportation, loading & unloading, installation etc. (complete as per Scope of Work) of quality control laboratory equipment along with complete accessories at SAPDC, Arun-3 HEP Dam site at Phaksinda, Sankhuwasabha Distt. Nepal” - Regarding clarifications.

Employer : SAPDC

Country : NEPAL

(This Document is meant for the exclusive purpose of bidding against this work and shall not be transferred, reproduced or otherwise used for purposes other than that for which it is specifically issued.)

Clarifications to the Bidders' Queries

Sr. No	Clause No.	Provision in Bid Documents	Summary of Bidders' Query / Modification requested	Response/ Comments
TERMS & CONDITIONS (T&C):				
1.	1 b 'Minimum Qualifying Requirements'	The firm/bidder should be manufacturer/authorized distributor/authorized supplier/ authorized reseller for 'Aimil' Brand. In addition to above, other suppliers can also participate in bidding subject to furnishing of a certificate from manufacturer/authorized distributor/authorized supplier/ authorized reseller for 'Aimil' Brand.	We gone through tender, they mentioned all the models and specifications are AIMIL make, do you want Aimil Make or Similar to Aimil?	Existing Provisions shall prevail.
2.	1 b 'Minimum Qualifying Requirements'	The firm/bidder should be manufacturer/authorized distributor/authorized supplier/ authorized reseller for 'Aimil' Brand. In addition to above, other suppliers can also participate in bidding subject to furnishing of a certificate from manufacturer/authorized distributor/authorized supplier/ authorized reseller for 'Aimil' Brand.	Against the above cited subject Tender we shall request you to kindly allow us to submit our offer to quote for our own 'ENKAY' Brand Equipment	Existing Provisions shall prevail.
"Annexure-B" TECHNICAL SPECIFICATIONS				
1.	-	-	In the Specification List it is given 27 items. But in the Bill of Quantity sheet it is mentioned 49 items.	The technical specifications mentioned under Annexure-B are for 27 Nos. items, however the technical specifications of remaining items are itself covered in the detailed nomenclature mentioned in the BOQ for supply of these items.
2.			In the specification sheet item 27 – Universal Testing Machine, the specification they mentioned is normal machine, but in the Bill of	With regard to the supply of UTM (BOQ item No 45), it is intimated that detailed

Sr. No	Clause No.	Provision in Bid Documents	Summary of Bidders' Query / Modification requested	Response/ Comments
			Quantity they mentioned 6 Pillar Hydraulic Gripping, please clarify which machine they required because price is huge differed.	technical specifications covering the required features have been mentioned. As such UTM computerised 1000KN capacity 6 pillar type with all accessories is required to be supplied as mentioned in the BOQ.
3.			Item 27 - Universal Testing Machine - specification they mentioned Extensometer Gauge 25-600mm, but in the Bill of Quantity they mentioned Extensometer Gauge 25 & 50mm. Please clarify, because both extensometer are different and while quoting price is also higher side.	The bidder/supplier shall supply the electronic extensometer as mentioned in BOQ (Sr No 45(i)) having the technical specification given for supply of the item.